

ADDENDUM Q & A

Jessica Lunardi
Fiscal Administrative Officer
847-412-0407
Telephone Number
jlunardi@mitchellplanning.com
E-mail Address
847-412-0385
Fax Number

UConn | HEALTH CENTER

STATE OF CONNECTICUT

UNIVERSITY OF CONNECTICUT HEALTH CENTER
Procurement Operations & Contracts
263 Farmington Avenue, MC4036
Farmington, CT 06032-4036

RFP NUMBER: 5-2624	PROPOSAL DUE DATE: March 27, 2015	PROPOSAL DUE TIME: 2:00 pm EST	RFP SURETY: N/A
RFP TITLE: Imaging Equipment- OR Hybrid			

ADDENDUM NUMBER: 1

DATE ADDENDUM ISSUED: March 20, 2015

FOR: The University of Connecticut Health Center

NOTE: To answer questions submitted by vendors,

This Addendum must be Signed & Returned with your proposal.

Authorized Signature of Proposer

Company Name

Approved By:



[Jessica Lunardi]

Fiscal Administrative Officer

(Original Signature on Document in Procurement Files)

** SEE ATTACHED Q & A **

QUESTION #1:	Specifications Tab: -Line 28 Dual Contrast. Please explain
UCHC RESPONSE:	Also known as dual energy.
QUESTION #2:	Line 35: Telescopic foot. Please explain
UCHC RESPONSE:	Also referred to as Telescopic pedestal
QUESTION #3:	Line 39: Table to be on vendor provided UPS (for table only) yet line 63 requests Complete system UPS
UCHC RESPONSE:	Vendor can provide a UPS for the table and the computer system or separate.
QUESTION #4:	Line 62. Option IVUS. Who would do that?
UCHC RESPONSE:	IVUS is becoming a useful tool in angiography. The quoted system needs to support the integration of the images on the system in-room display.
QUESTION #5:	RFP Scope Tab: -Berchtold is OEM of choice. Do they want the Large Display on the Berchtold boom or just lights, etc.
UCHC RESPONSE:	Display will need OEM boom.

END OF ADDENDUM